#### 8 Open Access Full Text Article

## ORIGINAL RESEARCH

Comparison of agreement and efficiency of a swept source-optical coherence tomography device and an optical low-coherence reflectometry device for biometry measurements during cataract evaluation

# Samuel F Passi\* Atalie C Thompson\*

Department of Ophthalmology, Duke University Eye Center, Durham,

\*These authors contributed equally to this work

Correspondence: Preeya K Gupta Department of Ophthalmology, Duke University Eye Center, 4709 Creekstone Drive, Suite 100, Durham, NC 27703, USA Tel +1 919 660 5234 Fax +1 919 660 5070 Email preeya.gupta@duke.edu



Preeya K Gupta NC, USA

Purpose: To compare the agreement and efficiency of a swept source-optical coherence tomography biometer, IOLMaster 700 (IOLM700), and a low-coherence optical reflectometry biometer, LENSTAR LS 900 (LS900), when acquiring biometry measurements during cataract evaluation.

Methods: A retrospective chart review of biometry measurements that were performed in 64 eyes of 32 patients on the same day using the LS900 and the IOLM700. The total image acquisition time per subject was compared between the two machines using a Wilcoxon ranksum test. Bland-Altman plots showing the mean difference and 95% limits of agreement were graphed. Intraclass correlation coefficients (ICCs) were calculated for the mean axial length (AL), anterior chamber depth (ACD), lens thickness (LT), and two keratometry mean values (K1 and K2) that were output from each device.

**Results:** The average time to complete biometry measurements in both eyes was significantly shorter for the IOLM700 compared with the LS900 (44.5±12.4 vs 168.8±67.2 seconds, P < 0.001). The Bland–Altman analysis and ICCs showed high degrees of agreement for the mean biometry values (ICC: AL 0.9999, ACD 0.9993, LT 0.9571, K1 0.9922, K2 0.9926) generated by the two devices.

**Conclusion:** There was a high level of agreement between the mean biometry output measures for IOLM700 and LS900. However, it took ~73% less time on average to acquire the images when using the IOLM700 compared with the LS900. In a busy clinic setting, use of the IOLM700 for biometry measures may save time and prove more efficient.

Keywords: IOLMaster, LENSTAR, biometer, cataract, intraocular lens, biometry

## Introduction

Biometry is an essential component of the cataract evaluation. Along with improved surgical techniques and fifth-generation intraocular lens (IOL) power calculation formulas, advances in biometric technology have allowed for unprecedented refractive outcomes.<sup>1</sup> Over the past few decades, ocular biometry has evolved rapidly away from the use of contact ultrasound probes to the implementation of noncontact imaging modalities. In 2000, the IOLMaster (Carl Zeiss Meditec AG, Jena, Germany) was released in the United States as the first noncontact automatic biometry device.<sup>2</sup> The initial IOLMaster employed partial coherence interferometry,<sup>3</sup> which utilizes a

Clinical Ophthalmology 2018:12 2245-2251

2245

Commercial Up of this work, is published and licensed by Dove Medical Press Limited. The full terms of this license are available at https://www.dovepress.com/terms.php
 and incorporate the Creative Commons Attribution — Non Commercial (unported, v3.0) License (http://creativecommons.org/license2/b-nd3.0/). By accessing the work you
 hereby accept the Terms. Non-commercial use of this work are permitted without any further permission from Dove Medical Press Limited, provided the work is properly attributed. For permission
 for commercial use of this work, please see paragraphs 4.2 and 5 of our Terms (https://www.dovepress.com/terms.php).

low-coherence light source to scan the eye while reflections are detected by a scanning mirror in order to produce optical A scans.<sup>4</sup> Ten years later, the IOLMaster 500 (IOLM500) was introduced and quickly gained popularity over the previous generation because it offered increased speed and dual-mode measurements that allowed for simultaneous axial length (AL) and keratometry calculations.<sup>2</sup>

The LENSTAR LS 900 (LS900; Haag-Streit AG, Koeniz, Switzerland) has become one of the most commonly used biometry devices since its Food and Drug Administration approval in 2009.<sup>5</sup> It combines optical low-coherence tomography with a 32-point pattern dual zone keratometry,<sup>6</sup> which contributes to the excellent level of precision delivered by the machine.<sup>7</sup> Utilizing optical low-coherence tomography, optical A-scans are acquired, and AL and anterior chamber depth (ACD) are calculated in a manner similar to the IOLM500. Moreover, the LS900 biometric parameters were expanded to include central corneal thickness (CCT) and lens thickness (LT).<sup>8</sup>

In 2015, the IOLMaster 700 (IOLM700) was released in the United States. This device has an integrated swept source-optical coherence tomography (SS-OCT) with a tunable laser of 1,055 nm as a light source, which generates optical B-scans allowing cross-sectional visualization of the eye. Because the narrow bandwidth light source reflections are projected to the eye one at a time, the SS-OCT has an improved signal-to-noise ratio.4 This enhanced visualization has improved the accuracy of biometry measurements in eyes with posterior subcapsular cataracts and dense cataracts.<sup>6,9–11</sup> Also, the IOLM700 measures CCT and LT,8 which was not possible with the IOLM500.8 The IOLM700 also utilizes telecentric keratometry, which combines a 950 nm light source and a unique optical configuration to ensure spot size remains constant irrespective of device-to-eye distance.<sup>6</sup> This technology makes patient movements less disruptive to image acquisition.

Although newer generation biometry devices strive to improve upon earlier versions,<sup>6</sup> it is not clear which biometry device is optimal. Moreover, there are many factors to consider when comparing biometry devices, ranging from accuracy and reproducibility to efficiency and cost. Previous studies have shown excellent correlation and agreement between the biometry output of the LS900 and IOLM500<sup>8,12</sup> and IOLM700.<sup>8</sup> However, to our knowledge, no study has yet compared the efficiency of the LS900 and IOLM700 in terms of the time they take to acquire images.

As patient volume has risen over the past decade, clinic efficiency has emerged as an increasingly important metric that physicians and clinic administrators assess.<sup>13,14</sup> The time

that it takes to acquire images during patient evaluation impacts clinic flow, which ultimately can affect the patient experience. Our study sought to compare two modern biometry devices, the IOLM700 and LS900, not only in terms of their level of agreement on biometry measurements but also on the total time they took to acquire the images.

## **Methods**

We conducted a retrospective review of adult subjects aged above 25 years who underwent evaluation for cataract surgery at a single academic center. This study was approved by the Institutional Review Board of Duke University Hospital and was conducted in accordance with the Declaration of Helsinki and the Health Insurance Portability and Accountability Act. A waiver of informed consent was granted due to the retrospective nature of this study. All data were de-identified for the purpose of statistical analysis.

All included subjects underwent a complete ophthalmic examination by a cornea specialist (PG). Eyes were excluded from the study if they had evidence of corneal or retinal pathology, prior refractive surgery, or abnormal structure, such as phacodonesis, white cataract, or corneal scar that would negatively impact the accuracy of biometry imaging. Subject eyes were imaged with both the IOLM700 and LS900 as part of their routine evaluation for cataract surgery. The order of the imaging test was random. Imaging technicians with extensive experience using both biometers recorded the start and stop time for total image acquisition for each subject by each machine as part of quality improvement.

The mean biometry values for AL, ACD, LT, and the flat (K1) and steep (K2) keratometry values were estimated by each machine (Figure 1). To compare the agreement between the two devices, two-way mixed effects models were used to calculate the intraclass correlation coefficient (ICC) for each of the biometry values. Linear regression analysis comparing the biometry values of the two machines was performed with clustered standard errors to account for multiple eyes within subjects, and the fitted values were graphed. Agreement between the two devices was also evaluated for each biometry value using the method of Bland and Altman, by plotting the difference between the measurements on the *v*-axis against the average of the measurements on the x-axis.<sup>15</sup> The 95% limits of agreements (LoA) were defined as the mean  $\pm$  1.96 SD of the difference between the measurements of the two devices. Total image acquisition time per subject was also compared between the two machines using a nonparametric Wilcoxon signed-rank test. A P-value < 0.05 was considered statistically significant.



Figure I Study flowchart.

Notes: Patients were imaged with both the IOLMaster 700 and the LENSTAR LS 900. The total acquisition time and biometry values for each device were recorded. Abbreviation: K, keratometry.

#### Results

Table 1 shows the demographic characteristics of the study subjects. A total of 64 eyes of 32 subjects were included in this study. The acquisition times for the IOLM700 and LS900 are compared in Table 2. The mean acquisition time was ~73% less for the IOLM700 (44.5 $\pm$ 12.4 seconds) compared with the LS900 (168.8 $\pm$ 67.2 seconds, *P*<0.001).

Table 3 shows the mean biometry values for AL, ACD, LT, K1, and K2 as well as the ICC. An ICC of <0.50 signifies poor agreement between the two devices, while 0.50-0.75 signifies moderate agreement, 0.76-0.90 good agreement, and 0.91-1.00 excellent agreement.<sup>16</sup> For all mean biometry values, there was excellent agreement between the two devices with all values having an ICC >0.90. Figure 2 shows graphs of the predicted values of the biometry measurements from regression analysis. Figure 3 shows the Bland–Altman plots for each of the biometry values with the mean difference and the upper and lower 95% LoA graphed. There were relatively few outliers: ACD had three, AL had four, LT had four, K1 had three, and K2 had one outlier. No definite systematic bias was observed.

## Discussion

Optical biometry is a key component of preoperative planning for cataract surgery. Accurate and reproducible biometry

Table		Demographics	of	the	study	subjects
-------	--	--------------	----	-----	-------	----------

67±10		
69.5 (35–83)		
12 (38)		
20 (62)		
25 (78)		
4 (13)		
3 (9)		

Abbreviation: N, number of patients.

measurements are critical to obtain the highest refractive accuracy and minimize postoperative refractive surprise. However, the time it takes to obtain measurements also deserves attention as it has the potential to impact clinic flow and patient experience. Our study confirms that the biometry values estimated by the IOLM700 achieve a high level of agreement with the LS900. However, to our knowledge, we are the first to show that image acquisition takes ~73% less time on average when utilizing the IOLM700 compared with the LS900. Our findings suggest that the IOLM700 may be more efficient than the LS900 for biometry measurement during routine cataract evaluation.

Several prior studies have sought to compare biometry devices in terms of their accuracy and their relative degree of agreement. Hui and Yi showed that the LS900 and IOLMaster V5.4 biometry values were highly correlated and that the IOL power calculations for the AcrySof SA60AT IOL were similar.<sup>17</sup> Both Rohrer et al<sup>12</sup> and Rasilber et al<sup>18</sup> have demonstrated very high to excellent correlation between the mean outputs of the IOLM500 and the LS900. In a 2017 metaanalysis by Huang et al that compared 18 studies involving 1,921 eyes, no difference was found in AL, ACD, and keratometry readings between the LS900 and IOLMaster.<sup>19</sup> In a more recent study of 120 eyes, Kunert et al demonstrated excellent agreement between the IOLM700, IOLM500, and LS900.8 Similar studies by Kurian et al and Shammas et al comparing the IOLM700 and the LS900 have also shown both good reproducibility and agreement.<sup>6,20</sup> Our study also demonstrates that the IOLM700 and LS900 show excellent

 Table 2
 Mean acquisition times of the IOLMaster 700 and LENSTAR LS 900

	LENSTAR LS 900	IOLMaster 700	P-value
Acquisition time (seconds)			
$Mean\pmSD$	168.8±67.2	44.5±12.4	<0.001

Note: P-value calculated with Wilcoxon signed-rank test.

Biometry values	LENSTAR LS 900	IOLMaster 700	Intraclass correlation
	Mean $\pm$ SD	Mean $\pm$ SD	coefficient (95% CI)
Axial length	23.7±1.25	23.7±1.24	0.9999 (0.9998–0.9999)
Anterior chamber depth	3.16±0.38	3.14±0.38	0.9993 (0.9989–0.9996)
Lens thickness	4.52±0.67	4.66±0.71	0.9571 (0.9268–0.9750)
Flat meridian KI	43.7±1.90	43.7±1.93	0.9922 (0.9869–0.9954)
Steep meridian K2	44.8±1.90	44.8±1.92	0.9926 (0.9874–0.9956)

Abbreviations: ICCs, intraclass correlation coefficients; K, keratometry.



Figure 2 Linear regression of the IOLMaster 700 vs the LENSTAR LS 900. Notes: (A) Axial length, (B) anterior chamber depth, (C) lens thickness, (D) keratometry I, and (E) keratometry 2. Abbreviation: K, keratometry.



Figure 3 Bland-Altman plots showing the mean difference (dotted line) and 95% upper and lower limits of agreement (solid lines) for biometry measurements obtained by the IOLMaster 700 and LENSTAR LS 900.

Notes: (A) Axial length, (B) anterior chamber depth, (C) lens thickness, (D) keratometry I, and (E) keratometry 2. Abbreviation: K, keratometry.

agreement in terms of their biometry output, but we are the first to show that the IOLM700 requires significantly less time to acquire images.

The ability of the IOLM700 to obtain images more quickly may be attributable to its integration of both telecentric

keratometry and SS-OCT technology.<sup>4,9</sup> Specifically, the rapid cycle, tunable 1,055 nm laser source has been shown to give a better signal-to-noise ratio and has improved tissue penetration.<sup>4,9</sup> Also, the LS900 may require better cooperation from subjects since all images are acquired in a single

position with one release procedure. If it is difficult for a patient to remain still during image acquisition, then it is possible that the LS900 measurements may be more difficult to acquire, leading to longer total imaging time. By comparison, the IOLM700 acquires images via three different positions and release procedures<sup>17</sup> and thus might require a shorter attention span by the subjects for each imaging position.

Limitations of our study include the retrospective nature of this study and the modest sample size. Also, we excluded eyes that had significant corneal or retinal pathology, which may limit the generalizability of these results. Finally, given the retrospective nature of the study, we did not directly evaluate whether the shorter time for image acquisition impacted clinic flow or patient experience. Future studies should aim to better elucidate whether the IOLM700 performs significantly better than the LS900 in terms of patient experience, clinic flow, and cost-effectiveness.

## Conclusion

Our study confirms that there is a high level of agreement in the mean biometry values generated from images acquired by the IOLM700 and LS900. However, image acquisition by the IOLM700 takes approximately three-fourths less time than image acquisition by the LS900, making it substantially more efficient. When compared with biometry machines that use older technology, machines that incorporate SS-OCT technology may save time and have the potential to improve clinic flow and patient experience.

## Disclosure

Dr Gupta is a consultant to Aurea, Allergan, Alcon, BioTissue, Johnson and Johnson Vision, Kala, Novabay, Novartis, Ocular Science, Shire, Tear Lab, Tear Science, and Zeiss. Dr Thompson is a recipient of the Heed Fellowship. All authors have no other conflicts of interest in this work.

#### References

- Kane JX, van Heerden A, Atik A, Petsoglou C. Intraocular lens power formula accuracy: comparison of 7 formulas. *J Cataract Refract Surg.* 2016;42(10):1490–1500.
- Carl Zeiss Meditec Receives FDA Clearance for the IOLMaster 500 Biometer. (2010, November 8). Available from: http://www.marketwired.com/printer\_friendly?id=1348694. Accessed May 5, 2018.
- Optical Biometry Explained. Available from: https://cataract-community. zeiss.com/optical-biometry-explained. Accessed May 5, 2018.
- Srivannaboon S, Chirapapaisan C, Chonpimai P, Loket S. Clinical comparison of a new swept-source optical coherence tomography-based optical biometer and a time-domain optical coherence tomography-based optical biometer. J Cataract Refract Surg. 2015;41(10):2224–2232.

- Goldman DA, Dee Stephenson P. Which device do surgeons prefer for optical biometry? (2015, August 25). Available from: https:// www.healio.com/ophthalmology/cataract-surgery/news/print/ocularsurgery-news/%7B4edddbf9-9621-41ef-84e6-046cbb8c98e1%7D/ which-device-do-surgeons-prefer-for-optical-biometry. Accessed May 25, 2018.
- Kurian M, Negalur N, Das S, et al. Biometry with a new swept-source optical coherence tomography biometer: repeatability and agreement with an optical low-coherence reflectometry device. *J Cataract Refract Surg.* 2016;42(4):577–581.
- LENSTAR LS 900 Improving outcomes Haag-Streit Diagnostics. (2016, August). Available from: https://www.haag-streit.com/fileadmin/ Haag-Streit\_USA/Lenstar/Landing/papers/Brochure\_Lenstar\_eng.pdf. Accessed May 25, 2018.
- Kunert KS, Peter M, Blum M, et al. Repeatability and agreement in optical biometry of a new swept-source optical coherence tomographybased biometer versus partial coherence interferometry and optical lowcoherence reflectometry. J Cataract Refract Surg. 2016;42(1):76–83.
- Shajari M, Cremonese C, Petermann K, Singh P, Müller M, Kohnen T. Comparison of axial length, corneal curvature, and anterior chamber depth measurements of 2 recently introduced devices to a known biometer. *Am J Ophthalmol.* 2017;178:58–64.
- Hirnschall N, Varsits R, Doeller B, Findl O. Enhanced penetration for axial length measurement of eyes with dense cataracts using swept source optical coherence tomography: a consecutive observational study. *Ophthalmol Ther.* 2018;7(1):119–124.
- Akman A, Asena L, Güngör SG. Evaluation and comparison of the new swept source OCT-based IOLMaster 700 with the IOLMaster 500. *Br J Ophthalmol.* 2016;100(9):1201–1205.
- Rohrer K, Frueh BE, Wälti R, Clemetson IA, Tappeiner C, Goldblum D. Comparison and evaluation of ocular biometry using a new noncontact optical low-coherence reflectometer. *Ophthalmology*. 2009;116(11): 2087–2092.
- Bohman B, Dyrbye L, Sinsky C. Physician well-being: the reciprocity of practice efficiency, culture of wellness, and personal resilience. *NEJM Catalyst.* 2017. Available from: https://catalyst.nejm.org/ physician-well-being-efficiency-wellness-resilience/. Accessed May 25, 2018.
- Milstein A, Lee TH. Comparing physicians on efficiency. N Engl J Med. 2007;357(26):2649–2652.
- Bland JM, Altman DG. Statistical methods for assessing agreement between two methods of clinical measurement. *Lancet*. 1986;1(8476): 307–310.
- Koo TK, Li MY. A guideline of selecting and reporting intraclass correlation coefficients for reliability research. J Chiropr Med. 2016;15(2):155–163.
- Hui S, Yi L. Comparison of two optical biometers in intraocular lens power calculation. *Indian J Ophthalmol.* 2014;62(9):931–934.
- Rabsilber TM, Jepsen C, Auffarth GU, Holzer MP. Intraocular lens power calculation: clinical comparison of 2 optical biometry devices. *J Cataract Refract Surg.* 2010;36(2):230–234.
- Huang J, McAlinden C, Huang Y, et al. Meta-analysis of optical low-coherence reflectometry versus partial coherence interferometry biometry. *Sci Rep.* 2017;7:43414.
- Shammas HJ, Ortiz S, Shammas MC, Kim SH, Chong C. Biometry measurements using a new large-coherence-length swept-source optical coherence tomographer. *J Cataract Refract Surg.* 2016;42(1):50–61.

#### **Clinical Ophthalmology**

#### Publish your work in this journal

Clinical Ophthalmology is an international, peer-reviewed journal covering all subspecialties within ophthalmology. Key topics include: Optometry; Visual science; Pharmacology and drug therapy in eye diseases; Basic Sciences; Primary and Secondary eye care; Patient Safety and Quality of Care Improvements. This journal is indexed on

Submit your manuscript here: http://www.dovepress.com/clinical-ophthalmology-journal

PubMed Central and CAS, and is the official journal of The Society of Clinical Ophthalmology (SCO). The manuscript management system is completely online and includes a very quick and fair peer-review system, which is all easy to use. Visit http://www.dovepress.com/ testimonials.php to read real quotes from published authors.

**Dove**press